

NA-MIC National Alliance for Medical Image Computing http://www.na-mic.org


Automatic Segmentation of Traumatic Brain Injury MRI volumes using Atlas Based Classification and 3D Slicer

Andrei Irimia, Micah C Chambers, John D. Van Horn Laboratory of Neuro Imaging University of California, Los Angeles andrei.irimia@loni.ucla.edu

NA-MIC Tutorial Contest: Summer 2011



- acquire familiarity with several MR sequences commonly used for TBI imaging
- learn how to apply ABC to perform joint co-registration and automatic segmentation of TBI volumes
- acquire expertise on how to identify and characterize TBI pathology using various MRI sequences
- gain exposure to informed strategies for quantification of TBI-related edema or hemorrhage





- This tutorial assumes that you have already completed the following tutorials:
- Slicer 3 Visualization Tutorial by Sonia Pujol, available at
 - <u>http://www.slicer.org/slicerWiki/images/2/2e/Slicer3_DataLoadingAndVisualization_UCSF2010_SoniaPujol.pdf</u>
- Interactive Editor Tutorial by Sonia Pujol, available at
 - http://www.slicer.org/slicerWiki/images/6/69/InteractiveEditorTutorial_Slicer3.6-SoniaPujol.pdf



This tutorial requires the installation of the **Slicer3.6 release** and the tutorial dataset. They are available at the following locations:

Slicer3.6 download page

http://www.slicer.org/pages/Downloads/

Tutorial dataset: TBISegmentationData_TutorialContestSummer2011.zip


http://www.na-mic.org/Wiki/images/7/74/TBISegmentationData_TutorialContestSummer2011.zip



- This tutorial has been developed using the Windows 7 platform. It is compatible with the following platforms:
 - Windows XP, Windows 7
 - Linux 32, Linux 64
 - Mac/Darwin




- Clinical background
- Clinical workflow
 - installation of the ABC module
 - loading and exploration of TBI data
 - understanding common MR sequences for TBI
 - automatic segmentation of TBI using ABC
 - generation of 3D models
 - lesion segmentation and model generation
- Conclusions



- Every year, an estimated 1.7 million traumatic brain injury (TBI) cases occur in the United States alone, with an estimated 1.2 million emergency room visits and over 50,000 deaths.
- Over 5.3 million Americans currently require long-term daily assistance as a result of traumatic brain injury (TBI), which occurs to 1.7 million US residents every year and results in health care costs of over \$60 billion/year
- Despite many innovations, progress towards patient-tailored characterization of the structural and functional substrates associated with TBI-related neural and cognitive impairment remains dissatisfactory and the relationship between neurophysiological markers of cognitive dysfunction and TBI structural damage has not been acceptably elucidated
- 3D Slicer offers a powerful and unparalleled set of tools for the exploration and quantification of TBI




- After Slicer has been loaded, the Welcome window will appear.
- To install the ABC module:
 - 1 Click on View → Extension Manager
 - 2 Click on Next
 - 3 Select ABC from the list of modules
 - 4 Click on Next; you will be prompted to specify the local hard drive location for ABC
 - 5 After installation, you must select Restart 3D Slicer now for the changes to take effect





- After Slicer has been loaded, the Welcome window will appear.
- To load the TBI volumes associated with the case study:
 - 1 Click on File → Add Data
 - 2 Click on Add File(s)
 - 3 Navigate to the data folder and sort the volume files in descending order by size. This will ensure that all files are later processed in proper order.
 - 4 Click on Open
 - 5 In the Add Data dialog box, click on Apply



3D Slicer Version 3.6.4-beta
 File Edit View Window Help

CtrLO

Ctrl-A

Load Scene...

Import Scene.


Add Volume...

Download Sample Da Add Data...

National Alliance for Medical Image Computing http://www.na-mic.org




- 4 TBI data volumes are provided for this tutorial: MP-RAGE, FLAIR, GRE, and SWI.
- MP-RAGE: this is an anatomical volume acquired using a T1-weighted sequence, in which white matter has higher intensity than gray matter
 - To explore a TBI-related lesion in this patient:
 - 1 link the views using S
 - The lesion associated with open-head TBI becomes apparent as a hypo-intensity located in the parieto-occipital region (see circle to the right)






- To locate the lesion more accurately within the head:
 - 1 navigate to slice 179 in the sagittal view
 - 2 navigate to slice 79 in the coronal view
- The lesion and open head injuries are now apparent on all three views
- Similar exploratory navigation can reveal the lesion in FLAIR, GRE, and SWI volumes



National Alliance for Medical Image Computing http://www.na-mic.org





MP-RAGE T1

FLAIR


GRE T2

SWI

- Use of multiple image channels provides complementary information on pathologies
- T1 is a standard volume, with good anatomical resolution and high WM/GM contrast
- FLAIR is excellent for revealing CSF-perfused lesions as image hyper-intensities
- T2 GRE is useful for discerning areas associated with hemorrhages
- SWI is suitable for the detection of micro-bleeds and is superior to GRE in this respect




- To load ABC, click on:
 - 1 –Modules → All Modules → ABC
- The ABC interface appears. To specify input data:
 - 2 click Parameter Set → Create New Command Line Module
 - 3 click Input image 1 → MP-RAGE.nii
 - 4 click Input image 2 → FLAIR.nii
 - 5 click Input image 3 → GRE.nii
 - 6 click Input image 4 → SWI.nii






- 1 click on and locate the *adult-atlas* directory, which was included in the tutorial data set
- 2 select adult-atlas from its location
- 3 click OK to select the atlas template
- 4 click Output label image → Create New Volume
- 5 Click Bias corrected output image 1 \rightarrow Create New Volume
- 6 Click Bias corrected output image 2 → Create New Volume
- 7 Click Bias corrected output image 3 → Create New Volume
- 8 Click Bias corrected output image 4 \rightarrow Create New Volume




Bias corrected output image 1 e Create New Volume Bias corrected output image 2 e MP-RAGE.nii Bias corrected output image 3 e FLAIR.nii Bias corrected output image 4 e SWL.nii Dias corrected output image 5 e Rename Edt Properties Delete	5 Bias corrected output image 2 Bias corrected output image 3 Bias corrected output image 4 Bias corrected output image 5 Cancel App Cancel App Delete	Bias corrected output image 3 Bias corrected output image 4 Bias corrected output image 5 Bias corrected output image 5 Cancel Ar Bias corrected output image 1 Bias corrected output image 1 Bias corrected output image 2 Rename Edt Properties Delete	Bias corrected output image 4 e . Bias corrected output image 5 e . Cancel App	Create New Volume None MP-RAGE.nii FLAIR.nii GRE.nii GRE.nii
--	---	--	--	---



- 1 click on the Advanced Parameters arrow
- 2 enter 0 (off) for the number of fluid iterations
- 3 click Apply to run ABC on the dataset
- ABC will now co-register the FLAIR, GRE and SWI images to the MP-RAGE image, and perform the tissue classification
- Allow up to several hours for ABC execution
- ABC produces 3 tissue categories:
 - gray matter
 - white matter
 - CSF







Now that we have obtained the label map of GM, WM and CSF using ABC, we can generate the 3D models associated with each of these three tissue types.

- 1 From the Modules drop-down menu, select All Modules
- 2 From the second column, select Model Maker

This will display the interface of the *Model Maker* module, which allows one to create 3D models in Slicer.

3D Slicer Version 3.6.4-bet	a		
File Edit View Window	Help Feedback		
Modules: UBC (A	All Modules	4D Image	Fast Rigid registration
-		ABC (Atlas Based Classification)	FastMarchingSegmentation
3DSlicer	Color	ACPC Transform	FetchMI
Sushicer	Data	Add Images	FiberBundles
$\overline{\gamma}$	Editor	BRAINSDemonWarp	Fiducial Registration
	Fiducials	BRAINSFit	FiducialSeeding
Help & Acknowledgement	Measurements	BRAINSROIAuto - Foreground masking tool	Fiducials
ABC (Atlas Based Classific	Models	BRAINSResample	Gaussian Blur
	ROI	BRAINSVectorDemonWarping	Gradient Anisotropic Diffusion
	SlicerWelcome	BSpline to deformation field	GradientAnisotropicFilter
	Slices	Binarize Map	Grayscale Fill Hole
	Transforms	Cast Image	Grayscale Grind Peak
Input Images	VolumeRendering	-	
	Volumes	ChangeTracker	Grayscale Model Maker
	Wizards	CheckerBoard Filter	Histogram Matching
	Informatics	ClipModel	IA_FEMesh
	Registration	CollectFiducials	IGTRecorder
	Registration	Color	IGTToolSelector
	Segmentation Quantification	CongealingCLI	Image Label Combine
	Diffusion	Convert Fiducials to Labelmap	Joint Rician LMMSE Image Filter
	Diffusion	Create Single ROI file	Label Map Smoothing
	IGT •	Create a DICOM Series	LabelStatistics
	Filtering	CropVolume	Labelmap Seeding
•	Surface Models	Curvature Anisotropic Diffusion	Linear registration
▲ Atlas	Converters •	Data	MRI Bias Field Correction
	Endoscopy	Dicom to Nrrd Converter	Mask Image
	Developer Tools	Diffusion Tensor Estimation	Measurements
	4D 🕨	Diffusion Tensor Scalar Measurements	Median Filter
Prior weight adjustments (sa	Diffusion Weighted 🕨	DiffusionWelcome	Merge Models
	Unspecified •	EMSegment Command-line	Mesh Contour Segmentation
 Output Images 	Utilities	EMSegmenter	Model Maker
L	Output label in	Editor	Model Transform
		Endoscopy	ModelIntoLabelVolume
Bi	as corrected output image	Execution Model Tour	ModelMirror
		Expert Automated Registration	Models
Bi	as corrected output image	Explode Volume Transform	Multiply Images
	-	Extract Baseline DWI Volume	N4ITK MRI Bias correction
Bi	as corrected output image	Extract Skeleton	NeuroNav
		Fast Affine registration	OpenIGTLink IF
Bi	as corrected output image	Fast Nonrigid BSpline registration	Orient Images
			Otsu Threshold




To generate the 3D models:

- 1 From the *Parameter Set* drop-down menu, select *Create New Command Line Module*
- 2 From the *Input Volume* drop-down menu, select *Output label image*
- 3 From the *Models* drop-down menu, select *Create New Model Hierarchy*

• 4 – Click Apply

These steps will create 3D models for WM, GM and CSF.








- 1 From the tool bar, click on the Models button
- 2 From the Set Model drop-down menu, select Model_1_jake
- 3 From the Set Model drop-down menu, select Rename and type WM; repeat steps 1-2 to rename Model_2_Peach as GM and Model_3_Brain as CSF
- 4 Set the model to WM as in step 2, and click on the color box;
- 5 In the Color Picker dialog box, click on the color drop down menu
- 5 Select Generic Anatomy Colors as the color map; this will assign a color scheme in accordance with Slicer conventions.








Now that we have selected an appropriate color scheme, we need to assign suitable colors to each tissue type.

 1 – In the Color Picker dialog box, select beige for WM


Repeat steps 2-5 on the previous slide for *GM* and *CSF*

- 2 For *GM*, click on purple as shown
- 3 For *CSF*, click on blue as shown. The 3D model should now look as to the right. Note the variable CSF thickness.



GenericAnatomyC(-








Now we can assign labels for the two pathology types, namely edema and bleeding:

- 1 From the label edit bar, click on *Level Tracing*
- 2 Click on the color button next to the Label
- 3 In the Color Box dialog box, select 31: edema





- 1 In the Foreground menu, select FLAIR.nii
- 2 Click on *the* tool bar menu
- 3 select *Red slice only*

Layout

 4 – Use the level tracing tool to segment the edema using the FLAIR volume.

2 🖳 🦾 🕻 🕻 🐼 Axial None MP-RAGE.nll-label ✓ MP-RAGE.nii ✓ I Conventional layout ଚ FLAIR.nii GRE nii Conventional widescreen layout SWI.nii Ø Four-up layout Output label image 88 Dual 3D layout Bias corrected output image 1 Bias corrected output image 2 Triple 3D layout Bias corrected output image 3 3D only layout Bias corrected output image 4 MP-RAGE.nii-label Red slice only layout 3 Rename... Yellow slice only layout Edit Properties... Green slice only layout Delete Tabbed 3D layout Tabbed slice layout Compare layout Compare widescreen layout Side-by-side lightbox layout

Toggle GUI panel visibility

Toggle GUI panel L/R

close

Recall that edema is hyperintense in this sequence modality, and that it appears in slices 105-117. The hypointense region in the middle is actually hemorrhage, but it will be segmented separately in following steps so we need not worry about it for now.

顀



Once the edema has been segmented, we can proceed to the segmentation of hemorrhage. Viewing the label map and the underlying image at the same time can be accomplished by adjusting label map opacity.

- 1 From the red slice image options menu, click on Adjust label map opacity
- 2 Use the slider to select a convenient opacity
- 3 In the Color Box dialog box, select 32: bleeding

This step ensures that 3D Slicer color labeling conventions are respected.



Label 1

Active Tool:


GREN


_ 🗎 🙋

_ \$

MP-RAGE nil-label

114 65.982








We can now segment bleeds.


• 1 – From the red slice background menu, click on *GRE.nii*

 2 – Use the label map opacity slider as shown on the previous slide as well as the level

tracing tool as demonstrated on slide 20 to obtain a convenient view of the hemorrhage in relation to the surrounding edema.

3 – Segment the bleeding in slices 106-116. The label map should now be similar to the inset shown to the right. Note that hemorrhage is surrounded by edemic region which does not bleed.





Create pathology 3D models

Once pathology has been segmented, their 3D models can be created.

- 1 From the tool bar menu, select the Conventional layout
- 2 For Input Volume, select MP-

RAGE.nii-label. This is the label map for the pathology.


• 3 – Under Models, select Create

New Model Hierarchy

• 4 – Under Model Maker Parameters,

type "31,32" in the Label input box.

• 5 – Click on Apply






Hierarchy & Display

Model Hierarch

We can now visualize the full model, which includes both pathology and healthy-looking tissue.

☆ & # \$

- 1 On the tool bar, click Models
- 2 Under Hierarchy and Display, select CSF




Hierarchy & Displa

5


Model Hierarchy

- 3 Under Visibility, uncheck the toggle box
- 4 Under *Hierarchy and Display, select GM*
- 5 Decrease the opacity to 0.2

This will allow us to visualize the edema and hemorrhage in relationship to the rest of the brain, as shown to the right.









1 – On the red box tool bar, toggle slice visibility (left below)

One can also display the pathology with GM/WM transparency (see figure to the right below)

• 2 – Under Hierarchy and Display, set WM opacity to 0.2






National Alliance for Medical Image Computing http://www.na-mic.org

Model Hierarchy		:
E- Scene		
WM		
- GM		
	- 10	
Hodel Maker Mo		
Model_31_ede		
Model_32_blee	aing	
Model Display		
Set Model:	GM	▲ ▼
	V	
Visibility		
-	.2	
-		



Segmentation of remaining lesions and hemorrhages in this subject can be accomplished using the tools and workflows demonstrated in this tutorial. Sample images of this undertaking are illustrated below.





- 3D Slicer offers powerful methodologies for the visualization of pathology due to traumatic brain injury
- Use of multiple MR image channels greatly enhances the ability to study and understand TBI structure/extent
- ABC is a robust algorithm to perform joint coregistration and automatic segmentation of TBI
- Completion of this tutorial allows one to acquire useful expertise on how to identify and characterize TBI
- Use of 3D Slicer can offer informed strategies for quantification of TBI-related edema or hemorrhage and for improved insight of clinical relevance





National Alliance for Medical Image Computing

NIH U54EB005149 (PI: Ron Kikinis MD; TBI sub-award: Jack Van Horn PhD)



National Institutes of Health

National Institute of Neurological Disorders and Stroke

ATIONAL INSTITUTE OF EUROLOGICAL ISORDERS AND STROKE NIH P01NS058489 (PI: Paul Vespa MD, FAAN, FCCN)



Laboratory of Neuro Imaging

University of California, Los Angeles

Director: Arthur W. Toga PhD



- A Irimia, MC Chambers, JR Alger, M Filippou, MW Prastawa, B Wang, DA Hovda, G Gerig, AW Toga, R Kikinis, PM Vespa, JD van Horn (2011) Comparison of acute and chronic traumatic brain injury using semi-automatic multimodal segmentation of MR volumes *submitted*
- A Irimia, MC Chambers, PM Vespa, AW Toga, JD van Horn (2011) Cortical network visualization and analysis in traumatic brain injury using multimodal neuroimaging *Proceedings of the 17th Joint Symposium on Neural Computation, Institute of Neural Computation, University of California, San Diego*
- A Irimia, MC Chambers, M Filippou, JR Alger, MW Prastawa, B Wang, S Gouttard, SMA Pujol, SR Aylward, DA Hovda, G Gerig, AW Toga, R Kikinis, PM Vespa, JD van Horn (2011) Threedimensional calculation and quantification of morphometric and volumetric cortical atrophy indices of widespread clinical use from MRI volumes of traumatic brain injury using 3D Slicer *Proceedings of the 41st Annual Meeting of the Society for Neuroscience (SfN 2011), Washington, DC* (1st Prize, Fine Science Tools Contest, UCLA)
- A Irimia, JD van Horn, MC Chambers, MW Prastawa, S Gouttard, PM Vespa, DA Hovda, JR Algers, SMA Pujol, G Gerig, SR Aylward, AW Toga, R Kikinis (2011) Automatic multimodal MR image segmentation for the clinical assessment of traumatic brain injury in 3D Slicer *Proceedings of the 17th Annual Meeting of the Organization on Human Brain Mapping (OHBM 2011), Quebec City, Canada*